

LAr R&D Progress Updates

Yichen, Milind, Aleksey, Steve

3/4/25



Lab Safety and Space Management

▶ **LN2 bulk order**

- 5000 gallon LN2 delivered last Friday
- Early delivery before 7am from CT

▶ **Bldg 510 power outage**

- Impact for multiple building
- I have requested for 1-month notification in advance
- Short 1 day shutdown is planned during the summer
- 5 day long term shutdown still expected later this year


▶ **HighBay AC work**

- Still in progress with power panel on LOTO
- Robotic testing stand is not running at this moment
- Air handler replacement expected tomorrow.
- A brand new unit will be installed


Diamond Substrate Emission Measurement Preliminary

► Increase the light input to the spectrometer

- Instead of free coupling, need to feed the light to spectrometer with fibers
- Received some collimators from Aleksey
- They have some issues in terms of connectivity and wavelength range and etc.
 - No improvement on the signal observed yet
- The substrate is 1" inch, ideally to have a 1" UV collimator with UV fiber
- Thomas has a 1" UV collimator but can't find it yet, a 1/4" UV collimator received
- Bob just provided a UV fiber to work with the UV collimator
 - Will make more attempts with the diamond substrate spectrum measurement



1 inch collimator



1/4 collimator






1/8 UV collimator

Diamond Substrate Emission Measurement Preliminary

► Wavelength range testing



- Spectrum measured for room light and Deuterium Lamp
- 1 inch collimator ruled out: No signal into the spectrometer due to non SMA connector
- 1/4 Collimator has a cut-off at 300nm
- 1/8 UV collimator has the lowest cutoff with the UV fiber
- 1/8 UV collimator chosen for the single wavelength measurement




Diamond Substrate Emission Measurement Preliminary

► Fine tuning for the maximum light collection

- It is noticed the signal after 1/8 collimator is still not very high
- The collimator size matched with the light slit width at the top
- Adjust the collimator height to the middle of the light spot
- The improvement is very clear with the spectrometer saturated at full spectrum
- When selecting 266 nm, the peak height increased by about 100x



Full Spectrum






266 nm selection

Diamond Substrate Emission Measurement Preliminary

▶ Diamond Substrate Emission Spectrum Measurement




- Insert the diamond in front of the collimator
- Significantly signal amplitude improvement



Diamond Substrate Emission Measurement Preliminary

▶ Diamond Substrate Emission Spectrum Measurement

- Very weak emission spectrum observed around 500 nm
- We are also collecting the 266nm wavelength
- Future improvement on collection would help
- Will try with the 1/4 collimator with 300nm cutoff
- Meeting with Thomas at 3pm to discussion the measurement improvement



Spectrometer exit slit repair

▶ Exit slit removed from the setup

- Both pieces moving after disassembly
- Do see signature of aged grease by the high resistance on the shutter
- Bill is disassembly entire slit to remove the old grease and re-lub the parts

