


bHCAL Meeting — Neutron Calibration Update

Jan Vanek

University of New Hampshire

03/06/2026



OVERVIEW



- Updated neutron calibration
 - Now able to calibrate 1 GeV neutrons
- Calibration using protons
 - Using information from tracking
 - Study to use proton information for neutrons

MANUAL CALIBRATION METHODS

- Same method used for protons and neutrons
- Method 2
 - $E_{calib} = A(E_{EMCAL} + BE_{bHCAL})$
 - Plot $(E_{EMCAL} + BE_{bHCAL})/E_{par,MC}$
 - First find B for which the distribution above has the smallest σ/μ
 - A is set as $1/\text{mean}$ of the distribution with optimal B
- More advanced method(s) to be implemented



UNCALIBRATED ENERGY

- Sum of energy deposition in bECAL+bHCAL
 - (left) proton
 - (right) neutrons
- Similar energy deposition distributions for 2 GeV and 5 GeV
- Clear difference for 1 GeV




CALIBRATION DISTRIBUTIONS

- Method 2 calibration distributions
 - (left) proton
 - Using energy (momentum) from tracking
 - Need to investigate origin of shoulders at low energies
 - (right) neutrons
- Update:** Require to have both bECAL and bHCAL information for calibration distributions




CALIBRATION DISTRIBUTIONS

- Method 2 calibration distributions
 - (left) proton
 - Using energy (momentum) from tracking
 - Need to investigate origin of shoulders at low energies
 - (right) neutrons
- Update:** Require to have both bECAL and bHCAL information for calibration distributions





Jan Vanek, bHCAL meeting



03/06/2026



CALIBRATION PARAMETER A

- Parameter A in Method 2 calibration
 - (left) proton
 - Using energy (momentum) from tracking
 - (right) neutrons
- **Update:** Require to have both bECAL and bHCAL information for calibration distributions





CALIBRATION PARAMETER B

- Parameter B in Method 2 calibration
 - (left) proton
 - Using energy (momentum) from tracking
 - (right) neutrons
- Update:** Require to have both bECAL and bHCAL information for calibration distributions





CALIBRATED DISTRIBUTIONS

- Calibrated energy distributions
 - (left) proton
 - (right) neutrons
 - 1 GeV distribution re-binned due to large calibration factors
- **Update:** Require to have both bECAL and bHCAL information for calibration distributions





CALIBRATED DISTRIBUTIONS

- Calibrated energy distributions
 - (left) proton
 - (right) neutrons
 - Detail for 1 GeV to show distribution shape
- **Update:** Require to have both bECAL and bHCAL information for calibration distributions





CALIBRATION RESULT — ENERGY

- Calibrated energy distributions
 - (left) proton
 - (right) neutrons
- **Update:** Require to have both bECAL and bHCAL information for calibration distributions






CALIBRATION RESULT — RESOLUTION

- Calibrated energy distributions
 - (left) proton
 - (right) neutrons
- **Update:** Require to have both bECAL and bHCAL information for calibration distributions






CALIBRATION DISTRIBUTIONS – LOW E

- First look at calibration distributions for low energy neutrons
 - 100 MeV, 200 MeV, and 500 MeV
 - Different x -axis scale compared to higher energies
- Using bECAL+bHCAL information



CALIBRATION DISTRIBUTIONS – LOW E

- First look at calibration distributions for low energy neutrons
 - 100 MeV, 200 MeV, and 500 MeV
- Using **Imaging ECAL+bECAL+bHCAL** information
- Neutron calibration for energies below 1 GeV difficult



SUMMARY

- Implemented calibration framework of bECAL+bHCAL system for protons
 - Tracking appears to provide good resolution to be used for calibration of protons, similar to π^+
- Updated calibration procedure for neutrons
 - Require both bECAL and bHCAL information for calibration
 - Can calibrate 1 GeV neutrons with updated framework (using MC info only for now)
- Protons and neutrons appear to have similar response for higher energies
 - 2 GeV and 5 GeV
- Substantial difference in response for lower energies (1 GeV)
- Calibration appears difficult for neutrons below 1 GeV
 - From first look at low energy neutrons in official campaign (100 MeV, 200 MeV, and 500 MeV)

THANK YOU FOR ATTENTION